

DECLARATION of CONFORMITY

We, NIKON INSTRUMENTS EUROPE B.V.
SCHIPHOLWEG 321, 1171PL BADHOEVEDORP,
THE NETHERLANDS ,

declare under our sole responsibility that the product,

TYPE of PRODUCT: Inverted Microscope
MODEL NUMBER : Ti-S, Ti-S/L100, Ti-U and Ti-U/B ,

*with HUB Controller B, TI-HUBC/B

to which this declaration relates is in conformity
with the following standards or other normative documents,

EN61010-1:2001 EN61010-2-101:2002 EN591:2001
EN61326:1997+A1:1998+A2:2001+A3:2003
EN55011:1998+A1:1999+A2:2002 Class B
EN61000-4-2:1995+A1:1998+A2:2001, EN61000-4-3:2002+A1:2002,
EN61000-4-4:2004, EN61000-4-5:1995+A1:2001,
EN61000-4-6:1996+A1:2001, EN61000-4-8:1993+A1:2001,
EN61000-4-11:1994+A1:2001
EN61000-3-2:2000+A2:2005, EN61000-3-3:1995+A1:2001+A2:2005,

following the provisions of

IN VITRO DIAGNOSTIC MEDICAL DEVICE Directive (98/79/EC)
LOW VOLTAGE Directive (2006/95/EC)
EMC Directives (2004/108/EC) .

Manufacturer's name: NIKON CORPORATION
INSTRUMENTS COMPANY

Location of
the manufacturer: 471 NAGAODAI-MACHI,
SAKAE-KU, YOKOHAMA-SHI,
KANAGAWA 244-8533 , JAPAN

BADHOEVEDORP/
October 29, 2007

日下 裕文
H. KUSAKA, President